


Instruction for FIP Ab test



Sample fill line (black line) : 1 drop (10 μ l)

1


Sampling & loading



1 drop of sample
 •• Do not overload the sample !!

2


Absorption



Wait for complete absorption

3

Buffer addition



2 drops of buffer
 •• after complete absorption



4

Reading

00:05:00-00:10:00



Read the result between 5 ~ 10 min



5

Discard

00:10:00



Discard the used device
 In valid result after 10 min

